Sanua_® Tokyo Japan

DIGITAL MULTIMETER

PM3

APPLICATIONS AND FEATURES

This instrument is a pocket type digital multimeter designed for the measurement of low-voltage circuit. This is used at small communications equipments, home electric appliances, voltage of lump line, and measurement of various types of batteries. 8.5mm thick body with multi-function

- •4000 count
- 0.7% best accuracy
- Resistance measurement
- Capacitance measurement
 Not suitable for measurement of conditions
- *Not suitable for measurement of condensers with large leak current.
- Frequency measurement(AC sine wave only)
- Duty cycle measurement
- Data hold
- Relative value
- Auto power save(15min after the last operation)(cancelable)

SPECIFICATIONS

	Measuring range	Best accuracy	Resolution	Input impedance
DCV	400m/4/40/400/500V	±(0.7%+3)	0.1mV	
ACV	4/40/400/500V	±(2.3%+10)	0.001V	DCV:
Resistance	400/4k/40k/400k/4M/40MΩ	±(2.0%+5)	0.1Ω	10M∼100MΩ
Capacitance	5n/50n/500n/5μ/50μ/200μF	±(5.0%+10)	0.001nF	
Frequency	9.999/99.99/999.9/9.99k/60.00kHz	±(0.7%+5)	0.001Hz	ACV:
Duty cycle	0.1~99%			10M~11MΩ
Continuity	Buzzer sounds at less than $10{\sim}120\Omega$. Open voltage : approx. 0.4V			
Diode test	Open voltage : approx. 1.5V			

Measuring method	Δ-Σ method		
Display	4000 count		
Range selection	Auto range		
Over display	"OL" is displayed		
Polarity	Automatic selection(only " - " is displayed when negative voltage is inputted.)		
Battery low warning	Battery() mark lights when the internal battery's power is low.		
Sampling rate	Approx. 3 times/sec		
Operating temperature / humidity	'0~40°C, 80%RH(Max), No condensation		
Storage temperature / humidity	-10~50°C, below 70%RH, No condensation		
Environmental condition	Altitude 2000m or below, pollution degree II		
Power consumption	6mW(at DCV)		
Bandwidth	40∼400Hz		
Battery	Coin type lithium battery CR2032(3V) x 1		
Size / Mass	H108 x W56 x D11.5mm / 85g		
Safety standards	EN61010-1 CAT.II 500V		
Standard accessories included	Case holder(C-PM3), Instruction Manual		





A battery for monitoring has been installed prior to shipment from the factory. It may be discharged before the expiration of the described battery life. This battery is used to check the functions and performance of the product. Specifications and external appearance of the product described above may be revised for modification without prior notice.

sanwa

SANWA ELECTRIC INSTRUMENT CO., LTD.

Dempa Bldg, 4-4 Sotokanda 2-Chome, Chiyoda-Ku, Tokyo 101-0021 Japan Tel:+81-3-3251-0941 Fax:+81-3-3256-9740

www.sanwa-meter.co.jp

Distributed by

