



FLEXIBLE CLAMP METER KEW 2204R/2210R



KEW 2204R/2210R Specifications

| | | KEW2204R | KEW2210R |
|-----------------------------------|------|---|--|
| AC A (RMS) | | | |
| Ra | ange | 4.000/40.00/400.0A | 30.00/300.0/3000A |
| Accuracy Crest factor (CF) | | ±3%rdg±5dgt[45-500Hz] (At the center of the circle formed by the flexible sensor) | |
| | | Full scale CF<1.6, half scale<3.2 Effective input crest values are $\sqrt{2}$ times of the max values of each range. | |
| Conductor size | | <i>φ</i> 70mm max. | ϕ 150mm max. |
| Influence of Conductor position | | Additional $\pm 2\%$ (max.) depending on the distance from the center position | Additional $\pm 3\%$ (max.) depending on the distance from the center position |
| Overload protection | | 500A AC for 10 seconds | 5000A AC for 10 seconds |
| Applicable standards | | IEC 61010-1, IEC 61010-2-032 CAT IV 600V / CAT III 1000V Pollution degree 2 IEC 61326-1(EMC) , IEC 60529 IP40 , EN50581(RoHS) | |
| Operating temperature & humidity | | 0 - +50°C,less than 80% RH (without condensation) | |
| Storage temperature & humidity | | -10 - +60℃,less than 70% RH (without condensation) | |
| Power source | | R03 / LR03(AAA)(1.5V)×2 **Continuous measuring time: approx.120hours (Auto power off: approx.15 minutes) | |
| Dimensions | | 120(L)×70(W)×26(D): Display unit | 120(L)×70(W)×26(D): Display unit |
| Weight | | 200g approx. (including batteries) | 300g approx. (including batteries) |
| Included Accessories | | 9174 (Carrying case) LR03(AAA)×2, Insruction manual | |

Operating Instructions

To open this flex sensor, disconnect the joint by pulling to the right.

Confirm that the joint connector on the flex sensor is firmly. Clamp onto one conductor only to the illustration above.





















Please read the "Safety Warnings" in the instruction manual supplied with the instrument thoroughly and completely Safety Warnings: In the instruction manual supplied with the instruction manual suppl to operate the instrument on a correct power supply and voltage rating marked on each instrument.

For inquires or orders:



2-5-20, Nakane, Meguro-ku, Tokyo, 152-0031 Japan Phone:+81-3-3723-0131 Fax:+81-3-3723-0152 E-mail:info-eng@kew-ltd.co.jp

http://www.kew-ltd.co.jp