



An inspection certificate is supplied as standard.
Refer to page U-9 for details.

Dial Indicators

Comparison measuring instruments which ensure high quality, high accuracy and reliability.

SERIES 1 — Compact Type, Extra Small Diameter

- Compact dial indicators with bezel diameters of 31 or 36 mm for restricted-space applications in gaging jigs.



1911A-10



Balanced scale




Graduation: 0.01 mm,
Measuring range: 2.5 mm

1911A-10

Jeweled bearing



Balanced scale



Graduation: 0.02 mm,
Measuring range: 0.5 mm

1913A-10


Jeweled bearing



1003A



Balanced scale



Graduation: 0.01 mm,
Measuring range: 4 mm

1003A

Dial Indicators

Comparison measuring instruments which ensure high quality, high accuracy and reliability.



An inspection certificate is supplied as standard. Refer to page U-9 for details.

DIMENSIONS

ISO/JIS
Type

1911A-10/1913A-10

ANSI/AGD
Type

1921A-10/1923A-10/1925A-10

1003A

Unit: mm

Unit: in

Metric

Unit: mm

Order No.	A	B	C	D	E	F	G	H	I	Mass (g)	
										w/lug	Flat-back
1911A-10	15.5	42	31	13.2	19.2	15	11.5	30	5.9	56	52
1913A-10	15.5	39.5	31	13.2	19.2	15	9	30	5.9	56	52
1003A	18	40.3	36	13.5	15	9.5	12.8	32	6	61	58

Inch

Unit: in

Order No.	A	B	C	D	E	F	G	H	I	Mass (g)	
										w/lug	Flat-back
1921A-10	0.61	1.57	1.22	0.52	3/4	0.59	0.37	1.18	0.27	56	52
1923A-10	0.61	1.51	1.22	0.52	3/4	0.59	0.31	1.18	0.27	56	52
1925A-10	0.61	1.48	1.22	0.53	3/4	0.59	0.28	1.18	0.27	56	52

Note 1:

Note 2:

Note 3:

SPECIFICATIONS

Metric

ISO/JIS type

Order No.	w/lug	Flat-back	Graduation (mm)	Range (range/rev) (mm)	Maximum permissible error (MPE) (µm)						Dial reading	Measuring force (N)
					Indication error				Hysteresis	Repeat-ability		
					1/10 Rev	1/2 Rev	1 Rev	Measuring range				
1911A-10	1911AB-10	0.01	2.5 (1)	8	9	10	12	4	3	0-50-0	1.8 or less	
1913A-10	1913AB-10	0.002	0.5 (0.2)	2.5	4	5	6	2.5	1	0-100-0	1.8 or less	
1003A	1003AB	0.01	4 (1)	8	10	11	13	4	3	0-50-0	1.4 or less	

Inch

ANSI/AGD type

Order No.	w/lug	Flat-back	Graduation (in)	Range (range/rev) (in)	Accuracy (in)			Repeat-ability (in)	Dial reading	Measuring force (N)
					First 1 Rev/2.5 Rev/10 Rev		Retrace			
					First 1 Rev	2.5 Rev				
1921A-10	1921AB-10	0.001	0.1 (0.04)	±0.001	±0.001	—	0.0002	±0.0002	0-20-0	1.8 or less
1923A-10	1923AB-10	0.0005	0.05 (0.02)	±0.0005	±0.0005	—	0.00016	±0.0001	0-10-0	1.8 or less
1925A-10	1925AB-10	0.0001	0.025 (0.01)	±0.0002	±0.0002	—	0.0001	±0.00003	0-5-0	1.8 or less

Note:

Note: